

Therapy Mouse User Guide



- Congratulations on your purchase of the Therapy Mouse motion control system—a Gyration® powered component integral to the operation of your Mobility Research system.
- Package Contents:**
- Air Mouse® (AMP) Module
 - LRC Remote Control (plus Coin battery)
 - USB Receiver Dongle
 - AMP Module USB Charging Cable
 - Therapy Mouse User Guide

THERAPY MOUSE
powered by Gyration®
*Motion Sensitive
Wireless Pointing Device*
Windows and OS X Compatible



Setup Instructions

Your Therapy Mouse is a Windows® and OSX® compatible motion-sensitive pointing device. Setup takes just a few minutes. After reviewing your Therapy Mouse components, follow these three simple steps:

- 1. Prepare the LRC Remote Control**
 - Install the LRC battery
 - Confirm USB Receiver pairing
(See Step 1 on reverse side)
- 2. Prepare the AMP Module.**
 - Charge the AMP Module battery
 - Confirm USB Receiver pairing.
(See Step 2 on reverse side)
- 3. Test Therapy Mouse Operation**
 - Move the AMP Module in the air and confirm screen cursor motion.
 - Press the LRC Remote Control's *Right* and *Left* click buttons and confirm computer click functionality.

Your Therapy Mouse System

Components

LRC Remote Control



Air Mouse® (AMP) Module



USB Receiver Dongle



FCC & CE Statements

This equipment has been tested and found to comply with the limits for a class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna, and/or increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

In order to maintain compliance with FCC regulations, shielded cables must be used with this equipment. Operation with non-approved equipment or unshielded cables is likely to result in interference to radio and TV reception. The user is cautioned that changes and modifications made to the equipment without the approval of manufacturer could void the user's authority to operate this equipment.

Industry Canada
This device complies with Industry Canada licence-exempt RSS standard(s). Operation is subject to the following two conditions: (1) this device may not cause interference, and (2) this device must accept any interference, including interference that may cause undesired operation of the device.

Le présent appareil est conforme aux CNR d'Industrie Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes: (1) l'appareil ne doit pas produire de brouillage, et (2) l'utilisateur de l'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.

CE Declaration of Conformity
This equipment has been tested and found to comply with the limits of the European Council Directive 1999/5/EC. A complete Declaration of Conformity can be found at www.SMK-Link.com.

NOTICES & WARNINGS

- This product may contain small parts. Keep out of the reach of small children.
- Battery Hazards:
 - Do not attempt to re-charge the supplied LRC Remote Control coin battery, but replace when necessary.
 - Do not dispose of your LRC coin battery in fire. This may cause them to explode, causing harm.
 - Do not attempt to disassemble any Therapy Mouse component.

© 2015 Gyration® is a registered trademark of SMK-Link Electronics, Inc. Apple® and OS X® are registered trademarks of Apple Inc. Windows® is a registered trademark of Microsoft Corporation. All other trademarks and registered trademarks are properties of their respective owners.



Mobility Research, Inc.
PO Box 3141,
Tempe AZ, 85280 USA
(800) 332-9255
www.litegait.com

Technical Support
technicalsupport@litegait.com
tel: (800) 332-9255
fax: (480) 829-0737

[PLACEHOLDER FOR MOBILITY RESEARCH THERAPY MOUSE WARRANTY]

Step 1: Preparing the LRC Remote Control

Install the LRC Remote Control Battery

Remove the LRC battery door (on bottom), turning it counter-clockwise with a small coin until released. Insert the coin battery with the "+" side facing up, and re-seat the battery door.

Confirm pairing with the USB Receiver Dongle

Your LRC Remote Control has been factory pre-paired with its USB Receiver Dongle. Should re-pairing become necessary, follow the steps listed in the *Pairing Instructions* section of this guide.

LRC Button Assignments

Left Mouse click	
Right Mouse click	
Keyboard Esc	
Keyboard Enter	



The LRC 1/8" Jack Input: By connecting an external device to this port, left and right mouse clicks can be activated from the external device instead of from LRC keypad.

Step 2: Preparing the AMP Module

Charging the Battery

Your AMP Module is powered by a rechargeable battery. Before initial use (or when the Module's color LED begins to flash every 5 seconds), the battery must be charged. To charge the battery, follow the steps listed in the *Charging & Calibration* section of this guide.

Confirm pairing with the USB Receiver Dongle

Your AMP Module has been factory paired with its USB Receiver. To re-pair, see the *Pairing Instructions* section of this guide.

AMP Module Switch Settings

Position	Mode	Air Motion	Cursor Movement	Active LED
OFF	Power OFF	None	None	OFF
REV	Reverse Mode			Orange
ON	Normal Mode			Green

Pairing Instructions

Re-pairing the LRC Remote with the USB Receiver

1. Plug the USB Receiver Dongle into a USB port on your computer.
2. After the receiver is recognized by the computer, press and hold the Dongle connect button for 5 seconds. Upon release, its LED will begin flashing blue.

3. Press the LRC Remote's *Left* () and *Right* () click buttons simultaneously and hold for 5 seconds. Upon release, the Remote's LED will begin flashing green.

4. When LEDs on both the LRC Remote and USB Receiver have stopped flashing, pairing is complete.

Re-pairing the AMP Module with the USB Receiver

1. Plug the USB Receiver Dongle into a USB port on your computer.

2. Press the USB Receiver's connect button and hold for 5 seconds. Upon release, the Receiver's LED will begin flashing blue.

3. Press the connect button on top of AMP Module and hold for 5 seconds. Upon release, the Module's LED will begin flashing green.

4. When LEDs on both the AMP Module and USB Receiver have stopped flashing, pairing is complete.

Charging & Calibration

Charging the AMP Module Battery

Your AMP Module is powered by an internal rechargeable battery. Before initial use (or when the Module's color LED begins to flash every 5 seconds), the battery must be charged. Follow these steps:

1. Connect the smaller end of the USB Charging Cable to the micro USB port of the AMP Module; and the other to a USB port on your computer (or a USB wall AC adapter - *not included*).

2. While charging, the orange LED on top of AMP Module will blink. When fully charged, the LED will light up solid green.

3. Remove the cable from AMP Module. Its LED will go off. Your AMP Module is now fully charged and ready for use.

Re-calibrating the AMP Module

Your AMP Module is factory calibrated. Should re-calibration become necessary, follow these simple steps:

1. Set the AMP Module's slide switch to the OFF position. Holding down the AMP button, slide the switch back to the ON position.
2. Now, place the AMP Module on a flat surface to calibrate. When its green LED flashes 5 times, calibration is complete.